

# GASCHECK G GAS TABLE

version 17 May 2011

For indication only. If in doubt, please contact info@ionscience.com



Gas Name	Trade Name	Formula	Response Factor	Minimum sensitivity ppm	Minimum sensitivity cc/sec	Gas Group
Air						
GAS GROUP 1			1	65	1.0E-05	
GAS GROUP 2			2.5	100	1.5E-05	
GAS GROUP 3			4	200	3.1E-05	
GAS GROUP 4			8	305	4.7E-05	
GAS GROUP 5			16	700	1.1E-04	
Helium		He	1.000	<b>65</b>	1.0E-05	1
Hydrogen		H <sub>2</sub>	0.769	<b>50</b>	7.7E-06	1
Ammonia		NH <sub>3</sub>	2.385	<b>155</b>	2.4E-05	2
Butane		C <sub>4</sub> H <sub>10</sub>	2.385	<b>155</b>	2.4E-05	2
Krypton		Kr	2.518	164	2.5E-05	2
Methane		CH <sub>4</sub>	2.903	189	2.9E-05	2
Neon		Ne	2.221	144	2.2E-05	2
Sulfur dioxide		SO <sub>2</sub>	2.647	172	2.6E-05	2
Sulfur hexa fluoride		SF <sub>6</sub>	2.231	145	2.2E-05	2
Trichloromethane		CHCl <sub>3</sub>	2.449	159	2.4E-05	2
1,1,2-Trichlorotrifluoroethane	R113	C <sub>2</sub> Cl <sub>3</sub> F <sub>3</sub>	2.600	169	2.6E-05	2
1,2-Dichlorotetrafluoroethane	R114	C <sub>2</sub> Cl <sub>2</sub> F <sub>4</sub>	2.750	179	2.7E-05	2
Dichlorodifluoromethane	R12	CCl <sub>2</sub> F <sub>2</sub>	2.697	175	2.7E-05	2
Bromotrifluoromethane	R1301	CBrF <sub>3</sub>	2.443	159	2.4E-05	2
Chlorodifluoromethane	R22	CHF <sub>2</sub> Cl	2.561	166	2.6E-05	2
refrigerant R 502	R502	CHClF <sub>2</sub> , CClF <sub>2</sub> HCF <sub>3</sub>	3.015	196	3.0E-05	2
Xenon		Xe	2.263	147	2.3E-05	2
Acetone		C <sub>3</sub> H <sub>6</sub> O	3.188	207	3.2E-05	3
Argon		Ar	3.527	229	3.5E-05	3
refrigerant R 404a	R404a	R125:143a:134a = 44:52:4	3.262	212	3.2E-05	3
refrigerant R 407c	R407c	R134a: R125: R32 = 40:40:20	3.277	213	3.3E-05	3
refrigerant R 410a	R410a	R125:R32 = 50:50	3.262	212	3.2E-05	3
refrigerant R 507	R507	CF <sub>3</sub> CH <sub>3</sub> :CF <sub>3</sub> CHF <sub>2</sub> = 50:50	3.846	250	3.8E-05	3
refrigerant R 245FA	R245FA	CF <sub>3</sub> CH <sub>2</sub> CHF <sub>2</sub>	3.600	234	3.6E-05	3
Boron trifluoride		BF <sub>3</sub>	3.912	254	3.9E-05	3
Carbon dioxide		CO <sub>2</sub>	4.029	262	4.0E-05	3
Deuterium oxide		D <sub>2</sub> O	4.597	299	4.6E-05	3
Diethyl ether		C <sub>4</sub> H <sub>10</sub> O	4.005	260	4.0E-05	3
Ethanol		C <sub>2</sub> H <sub>5</sub> OH	4.213	274	4.2E-05	3
Hexane		C <sub>6</sub> H <sub>14</sub>	3.667	238	3.7E-05	3
Hydrogen chloride		HCL	3.106	202	3.1E-05	3
Hydrogen sulphide		H <sub>2</sub> S	3.225	210	3.2E-05	3
Methanol		CH <sub>3</sub> O	4.330	281	4.3E-05	3
Nitrous oxide		N <sub>2</sub> O	4.271	278	4.3E-05	3
Pentane		C <sub>5</sub> H <sub>12</sub>	3.981	259	4.0E-05	3
Perfluorocyclobutane	C318	C <sub>4</sub> F <sub>8</sub>	3.106	202	3.1E-05	3
Tetra fluoromethane	R14	CF <sub>4</sub>	3.804	247	3.8E-05	3
Trichlorofluoromethane	R11	CFCl <sub>3</sub>	3.164	206	3.2E-05	3
Water		H <sub>2</sub> O	4.634	301	4.6E-05	3
Acetylene		C <sub>2</sub> H <sub>2</sub>	7.692	500	7.7E-05	4
Ethane		C <sub>2</sub> H <sub>6</sub>	7.962	518	7.9E-05	4
Ethylene Oxide		C <sub>2</sub> H <sub>4</sub> O	7.692	500	7.7E-05	4
Ethylene		C <sub>2</sub> H <sub>4</sub>	10.120	658	1.0E-04	4
Isobutane	R600a	C <sub>4</sub> H <sub>10</sub>	4.965	323	4.9E-05	4
Propane		C <sub>3</sub> H <sub>8</sub>	7.022	456	7.0E-05	4
Tetrafluoroethane	R134a	C <sub>2</sub> H <sub>2</sub> F <sub>4</sub>	5.846	380	5.8E-05	4
Carbon monoxide		CO	11.538	750	1.2E-04	5
Nitric oxide		NO	25.801	1677	2.6E-04	5
Nitrogen		N <sub>2</sub>	11.538	750	1.2E-04	5
Oxygen		O <sub>2</sub>	28.846	1875	2.9E-04	5